

D Brash & Sons Ltd

37 Stamperland Crescent Clarkston, Glasgow, G76 8LH
 sales@brash-scales.com www.brash-scales.com



No: 183221

Certificate of Calibration

Crossways Recycling 15 Manor Way Business Park Swanscombe DA10 0PP

Customer Ref:**Location:****Manufacturer:** Dini Argeo**Type / Model:** DFWL**Serial No:** 0100594170**Capacity:** 50000 Kg x 20 Kg**Expires end:** October 2023

All test equipment used is certified and traceable to the National Standards

Test Equipment Details

Set Number	Certificate Number
C1 to C26	9781

Maximum Permissible Error(MPE) +/-

0 Kg - 10000 Kg	+/- 20 Kg
10000 Kg - 40000 Kg	+/- 40 Kg
40000 Kg - 50000 Kg	+/- 60 Kg

AS Found Hysteresis

Test Point Kg	Reading Kg	Error Kg
0	0	0
2000	2000	0
10000	9980	-20
20000	19960	-40
31560	31500	-60
43560	43480	-80

Post Adjustment Hysteresis

Test Point Kg	Reading Kg	Error Kg
0	0	0
2000	2000	0
10000	10000	0
20000	20000	0
31560	31540	-20
43560	43520	-40

As Found Repeatability

Zero Kg	Load Kg	Zero Kg	Load Kg	Zero Kg	Load Kg
0	43520	0	43520	0	43500

Post Adjustment Repeatability

Zero Kg	Load Kg	Zero Kg	Load Kg	Zero Kg	Load Kg
0	43520	0	43520	0	43500

As Found Eccentric Loading

Reading Kg	Error Kg	Reading Kg	Error Kg
1 7020	20	2 7020	20
3 6980	-20	4 6980	-20
5 6980	-20	6 6960	-40
7		8	

Post Adjustment Eccentric Loading

Reading Kg	Error Kg	Reading Kg	Error Kg
1 7020	20	2 7020	20
3 6980	-20	4 7000	0
5 7000	0	6 6960	-40
7		8	

Technicians Comments

Weighbridge Tested & Calibrated As Necessary

This equipment has been tested and confirmed to be within the tolerances detailed above by the application of testing procedures for non-automatic weighing equipment.

All stated tolerances are for guidance only. Tolerances vary based on age, condition and the technical limitations of the equipment being tested they are calculated allowing for the MPE of test equipment, environmental, conditions and OEM in-service tolerances if available.

Engineers Name: **Adrian Jones**Date of Issue: **Mon, 31 Oct 2022**